

Government of West Bengal.
Office of the Principal, Burdwan Medical College.
Purba Bardhaman.

Memo No. BMC/PC/

Dated, Purba Bardhaman, the

2023.

QUOTATION

Sealed quotation are hereby invited from the Manufacturer Company / Dealers or Suppliers, for supply the undernoted instrument to **NPSVH project** related to the **Microbiology Department**, Burdwan Medical College, Purba Bardhaman. Rates should be inclusive of taxes, delivery charges and all other charges if any, failing which the quotation will be treated as cancelled. Quotation should be reached this office on or before 26/05/2023 within 1PM in a sealed cover mentioning the purpose of Quotation. No quotation will be received after the stipulated date and time. The undersigned reserves the right to accept or reject any quotation without assigning any reason. Valid trade license, P.Tax & Pan Card (Xerox Copy) GST Registration and others terms and condition which mentioned in other enclosure, should be submitted with the rate Quotation.

Sl no.	Name of the Articles.	Rate Per
1.	Vortex (Specification attached here with)	Per pc

Principal,
Burdwan Medical College.
Burdwan

Memo No. BMC/PC/ 1309/1(7)

Dated, Purba Bardhaman, the 19/5/ 2023.

Copy forwarded to:-

1. The District Magistrate, Purba Bardhaman.
2. The Information and cultural officer, Purba Bardhaman is being requested to upload the Quotation notice in District NIC Website.
3. The Chief Medical Officer of Health, Purba Bardhaman.
4. The M.S.V.P., B.M.C. & hospital, Purba Bardhaman.
5. www.wbhealth.gov.in, Swasthya Bhawan, Salt Lake, Kolkata-700091
6. Burdwan Medical College, Website, Purba Bardhaman.
7. College Notice board, for favour of information with the request to please display the notice in the Notice Board for wide circulation.

19/5/23
Principal,
Burdwan Medical College.
Burdwan

■ VORTEX MIXER

1. Sturdy housing ,metal coated,disinfectable
2. Digital display of speed
3. Rotation speed 0-2500rpm, strong orbital movements
4. Standing on four rubber feet design to prevent sliding of the instrument during shaking
5. Two operating modes : continuous and touch function
6. Slip resistant stand
7. Overheating protection
8. Permissible ambient temperature and relative humidity: 5-40°C and $\leq 80\%$
9. Should provide integrated rubber adapter with tube holder and one additional rubber adapter.
10. Adapter for mixing up to 18 micro tubes (6 x 0.5 ml, 6 x 1.5 ml).
11. Should have speed regulator control system.
12. It should have the interchangeable platform pad with variable for 50 ml tube and micro centrifuge tube.
13. Should have heavy duty AC motor.
14. Must have unique touch feature operate the unit when tube is pressed on the rubber cup.

General specifications:

1. Satisfactory working report of the quoted model from three reputed government institutes/hospitals should be included(within two years of submission of bid).
2. All equipments should specify Design qualifications, Installation qualifications, Operational Qualifications and Performance qualifications. Validation and calibration reports should have traceability towards applicable national / international standards.
3. Original product catalogue should be provided by the firm.Necessary documents, technical write up in English shall be attached with the offer both in hard and soft copies. Performance, efficiency, other factors such as distortion etc. as applicable be also furnished.
4. Installation: the bidder must arrange for the equipment to be installed by certified or qualified personnel; any prerequisites for installation to be communicated to the laboratory in advance. Onsite, in lab customer training to be provided.The equipment should be supplied with 2 meter cord at input and fitted with plugs of appropriate rating(5/15Amp)
5. 5 years back to back warranty followed by 5 year CMC should be provided by the company and warranty should commence only after delivery of satisfactory working report.
 - a. Complete with comprehensive set of spare parts.
 - b. The make, rating, model, description, specifications, quantity of each item shall be furnished separately.

S.

Yam
Chab

14

Yam

Yam

Yam

WJ

Yam

- c. Downtime commitment to be restricted to maximum period of 48 hours in metro cities and 96 hours in hard to reach areas (working days) and punitive actions shall be taken in case of failure to maintain the desirable downtime.
- d. Preventive maintenance plan and technical support to be provided.

6. Certifications:

- a. Product Certification : CE Class II A or US FDA certified
- b. Quality Certification : BIS/ISO certified
- c. Electrical Safety : Equipment meets electrical safety specifications such as that of IEC (Class I)

The committee approved the specification of Vortex mixer

Man
Vans *Chait* *8.* *9.* *us*

Subre *[Signature]*